

National Institute  
of Standards and Technology



National Voluntary  
Laboratory Accreditation Program

ISO/IEC 17025:1999  
ISO 9002:1994

## Scope of Accreditation



Page 1 of 1

### CALIBRATION LABORATORIES

NVLAP LAB CODE 200536-0

ASANUMA GIKEN CO., LTD.

4079-1 Koto-Cho, Hamamatsu-shi

Shizuoka-ken 431-1103

JAPAN

Mr. Susumu Asanuma

Phone: 81-53-486-1110 Fax: 81-53-486-1109

E-Mail: papa@asanuma-g.jp

### DIMENSIONAL

NVLAP Code: 20/D16

Coordinate Measuring Machines

#### Range

0 to 150 mm

#### Best Uncertainty ( $\pm$ )<sup>note 1</sup>

3.0  $\mu$ m

#### Remarks<sup>note 2</sup>

Artifact Calibration

1. Represents an expanded uncertainty using a coverage factor,  $k=2$ .
2. Calibration of Asanuma Giken manufactured Quality Master Gauge, two and three dimensional measurements.

March 31, 2004

Effective through

A handwritten signature in cursive script, appearing to read "C. L. Lavin".

For the National Institute of Standards and Technology